

Issued by NMI Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands

In accordance with The Council Directive 2009/23/EC on non-automatic weighing instruments.

Manufacturer SC Cantarprod S.r.L.  
Street Dambovita 12  
900309 Constanta  
Romania

In respect of A class **(III)** or **(III)**, electronic, single- or multi-interval or multi-range  
**non-automatic weighing instrument.**  
Manufacturer mark/name: SC Cantarprod S.r.L.  
Type : Depending on the configuration

Characteristics  $n \leq$  the number of scale intervals mentioned in the test certificates involved  
In the description number T8036 revision 0 further characteristics are described.

Valid until 28 October 2021

Description The instrument is described in the description number T8036 revision 0.

The Notified Body no. 0122  
NMI Certin, 28 October 2011



C. Oosterman  
Head Certification Board



# Description

Number **T8036** revision 0  
Project number SO 11200962  
Page 2 of 3

## 1 General information about the non-automatic weighing instrument

All properties of the non-automatic weighing instrument, whether mentioned or not, may not be in conflict with the legislation.

### 1.1 Essential parts

Indicator:

Manufacturer	Type	Test certificate number
Rinstrum	R 3xx series	TC6242
Rinstrum	R 4xx series	TC6821
Rinstrum	X 3xx series	TC7419

Any load cell(s) may be used under this EC type-approval certificate for instruments as described in WELMEC 2.4 Issue 2, provided the following conditions are met:

- There is a respective test certificate (EN45501) or an OIML Certificate of Conformity (R60) issued for the load cell by a Notified Body responsible for type examination under Directive 2009/23/EC.
- The certificate contains the load cell types and the necessary load cell data required for the manufacturer's declaration of compatibility of modules (WELMEC 2 Issue 5 Section 11), and any particular installation requirements. A load cell marked **NH** is allowed only if humidity testing to EN45501 has been conducted on this load cell.
- The compatibility of load cells and indicator is established by the manufacturer by means of the compatibility of modules form, contained in the above WELMEC 2 document, at the time of EC verification or declaration of EC conformity of type.
- The load transmission must conform to one of the examples shown in WELMEC 2.4 Issue 2.

### 1.2 Essential characteristics

The essential characteristics are described in the test certificates involved.

### 1.3 Essential shapes

The data plate is secured against removal by sealing or will be destroyed when removed. To secure components that may not be dismantled or adjusted by the user, the non-automatic weighing instrument has to be secured in a suitable manner on the locations indicated in the test certificates involved and the load cell cable or the junction box has to be secured.



# Description

Number **T8036** revision 0  
Project number SO 11200962  
Page 3 of 3

## 1.4 Conditional parts

The non-automatic weighing instrument may be equipped with peripheral equipment which is used for the applications listed in article 1(2)(a) of Directive 2009/23/EC, if the peripheral equipment is certified to be connected to an EC type-approved non-automatic weighing instrument by a Notified Body appointed to certify non-automatic weighing instruments according to paragraph I of Annex II of the EC Directive on Non-Automatic Weighing Instruments.

The non-automatic weighing instrument is fitted with a levelling device and a level indicator, unless the instrument is installed in a fixed position. The level indicator has a sensitivity of at least 2 mm for a tilt of 2/1000.

## 1.5 Conditional characteristics

The conditional characteristics are described in the test certificates involved.

## 1.6 Non-essential parts

The non-automatic weighing instrument may be connected to non-essential devices, for example but not limited to bar code readers, foot switches, second display's and cash drawers, provided that:

- They do not present primary data used for purposes mentioned in article 1(2)(a) of Directive 2009/23/EC unless the "preliminary observations" in Annex 1 of this directive is satisfied;
- They do not lead to an instrument having other essential characteristics than those fixed by this type-approval document.

## 2 Approval conditions

See chapter 1.3, essential shapes.

## 3 Seals and verification marks

See chapter 1.3, essential shapes.

## 4 CE-mark of conformity and inscriptions

The marks, facilities for the marks and the inscriptions on the non-automatic weighing instrument fulfill the requirements of article 1 of Annex IV of Directive 2009/23/EC.